


# Application reference

Solaris® gold bath supra can be used in a Wieland AGC Micro Plus unit.


No activator is required.

The amount of Solaris gold bath supra required for precipitation can be determined from the corresponding table for the Wieland AGC Micro Plus unit.

All other device parameter settings can be determined from the corresponding original instruction manual of the unit.



Dr. Udo Schusser  
Director, Research and Development



Dr. Markus Vollmann  
Director, Application Technology

Last revision: Feb. 2006